



Republic of the Philippines
CAVITE STATE UNIVERSITY
(CvSU)
Don Severino De Las Alas Campus
Cavite, Philippines

www.cvsu.edu.ph

This Bid Bulletin is issued to modify or amend items in the Bid Documents. This shall form an integral part of the bid Documents. Please take note of the changes / adjustments for the project,
Supply, Delivery and Installa

- Protection Class According to DIN EN 60529: IP 20 or better
- Motorized Lift with Safety-Stop Function
- Timer Function
- Highly Efficient 1500 cm² Condenser Unit
- Vertical Glassware
- Bath Volume Max. [l]: at least 3 L
- Voltage [V]: 220 – 240

NOTE: Must fit inside the Fumehood

Lot 2

Item No. 1-	2 units Stainless Steel Counter Laboratory Bench	200,000.00	400,000.00
	Rack Dimension: 4ft x 12ft x 6ft		
	- 13 Cabinet/2 swing door/with adjustable shelves/14 drawers/2 legroom		
	- 2 laboratory stool, stainless steel (304) frame with fiber glass top seat		
Item No. 2-	2 units Emergency shower with eye wash	- 28,000.00	56,000.00
	Spray section		
	- Nozzle height: at least 220cm		
	-		

